

GOVERNMENT OF ASSAM
OFFICE OF THE DIRECTOR OF MEDICAL EDUCATION, ASSAM
SIXMILE, KHANAPARA, GUWAHATI-22
Tele-Fax No.0361-2366236, E-mail- dmeassam@gmail.com, dme@assam.gov.in,

No. DME/MRI/241/2012/ 1810,

Dated: 17-02-2014

CORRIGENDUM NOTICE

One change is made in the tender vide No. DME/MRI/241/2012/897, Dated 22-01-2014 for supply of "Whole Body 16 Slice C.T. Scanner in Page No. 37, column 9 of heading H. Patient table as:

"System should have Carbon Fibre indexed Flat Table Top for Radiotherapy Treatment planning as standard"

Other terms and conditions including last date of submission are remained same.


Sd/- Prof.D.Hazarika,
Director of Medical Education, Assam.

Memo No. DME/MRI/241/2012/ 1811-19

Dated: 17 - 02-2014

Copy for forwarded for information & necessary action to:

1. The Commissioner & Secretary to the Govt. of Assam, Health & F.W.(B) Deptt. Dispur, Guwahati.
2. The P.S. to the Hon'ble Minister, Health & F.W. Assam for kind appraisal of the Hon'ble Minister.
3. The office Website of this Directorate.


Director of Medical Education, Assam