

**GOVERNMENT OF ASSAM  
OFFICE OF THE DIRECTOR OF MEDICAL EDUCATION, ASSAM  
SIXMILE, KHANAPARA, GUWAHATI-22**

No. DME/35/2013/10498

Dated: 18 - 05-2013

**INVITATION FOR BIDS  
(NATIONAL COMPETITIVE BIDDING)  
IFB NO.DME/35/2013/9876 , Dated: 06- 05-2013  
CORRIGENDUM CUM ADDENDUM**

**Modified Technical Specifications**

**3. OPERATING MICROSCOPE**

<b>Name of the features</b>	<b>Original</b>	<b>Modified</b>
<b>Apochromatic Optics</b>	Motorized focus; working distance <b>200-50 mm</b> Motorized zoom; 1:6 zoom ratio 10 x magnetic wide field eyepiece with integrated eyecups 125 x magnetic wide field eye <b>Speed focus for automatic focusing (only with integrated video camera)</b>	<b>200-50 mm to be deleted, and it is to be specified by the bidders</b>  <b>Speed focus for automatic focusing (only with integrated video camera) is to be deleted</b>
<b>Illumination</b>	<b>Supelux 180w xenon light source</b> Supelux 300w xenon light source with automatic Lamp exchange Automatics Iris control for adjusting the illumination to the field of view Display of remaining lamp life time	<b>Supelux 180w xenon light source</b> To be deleted
<b>System Setup</b>	Manual balance with guidance via user interface Auto balance <b>Auto Drape – automatic air evacuation out of the drape</b>	<b>Auto Drape – automatic air evacuation out of the drape is deleted</b>
<b>General Video</b>	Integrated 17 digital display Video Speed focus Automated video focus Adaptation of consumer (SLR) Photo/Video Camera	<b>General Video &amp; Standard definition video the features are to be deleted.</b>
<b>Standard definition video</b>	Integrated 3 – CCD SD video camera Integrated image and video	
<b>High- definition video</b>	Integrated 3 – CCD SD video camera Integrated recording device	<b>Beam splitter with stereo co-observer to be inserted.</b> <b>With recording and editing facilities to be added.</b>
<b>Connectivity/Data management</b>	Network connection via shared directory Video-in fore external SD video sources	
<b>System Operation</b>	Multifunctional Programmable handgrips Magnetic cultch for all system axes Central user interface	

	Overhead display for showing important system parameters XY drive (variable speed) Wire foot control panel Wireless foot control panel	
<b>Accessories</b>	<b>180° tilt able tube focal length f-170mm</b> <b>180° tilt able tube focal length f-200mm</b> <b>Foldable Tube f 170/f 260 including PROMAG function for addition 50% magnification and integrated rotate function</b> <b>Straight tube, focal length f-170mm</b> Stereo co-observer Stereo bridge for face to face operation Interface for micromanipulator	<b>180° tilt able tube focal length f-170mm</b> <b>180° tilt able tube focal length f-200mm</b> <b>Foldable Tube f 170/f 260 including PROMAG function for addition 50% magnification and integrated rotate function</b> <b>Straight tube, focal length f-170mm</b> <b>The above are to be deleted.</b> <b>The followings are to be inserted-</b> <ol style="list-style-type: none"> <li>1. <b>HD Camera</b></li> <li>2. <b>Beams splitter</b></li> <li>3. <b>camera mount</b></li> <li>4. <b>2 nos. spare xenon bulb.</b></li> </ol>
<b>Technical Data</b>	Floor stand system Electrical data Rated Voltage 115 VAC (100-125 VAC) 230 VAC (220-240 VAC) Current Consumption 115 VAC max 10A 230 VAC max 5A Rated frequency 50-60 Hz <b>Weight</b> Total weight approx 275 Kg. Maximum admissible load (accessories) on microscope head may not exceed 7 Kg. <b>Weight of transport container Approx 220 Kg.</b>	<b>Weight of transport container Approx 220 Kg is to be deleted.</b>  <b>The unit should be easily transportable.</b>  <b>The unit should be CE/ U.S - FDA certified.</b>

The **Item No. 13, i.e. Crash Cart** of the Plastic Surgery Department is dropped from the Tender.

Sd/- Prof. D. Hazarika,  
Director of Medical Education, Assam

Memo No.DME/35/2013/10499-503

Dated: 18 / 05/2013

Copy for forwarded for information & necessary action to:

1. The Secretary to the Govt. of Assam, Health & F.W.(B) Deptt. Dispur, Guwahati.
2. The Mission Director, N.R.H.M, Assam. He is requested kindly to upload the CORRIGENDUM CUM ADDENDUM notice in the NRHM, Assam Website. Soft copy enclosed.
3. The State Informatics Officer, Block F, Assam Secretariat, Dispur, Guwahati-06. He is requested to upload the CORRIGENDUM CUM ADDENDUM Notice in the Assam Govt. Website. Soft copy enclosed.
4. The P.S. to the Hon'ble Minister, Health & F.W. Assam for kind appraisal of the Hon'ble Minister.
5. The Notice Board of this Directorate.

Sd/- Prof. D. Hazarika,  
Director of Medical Education, Assam