

GOVERNMENT OF ASSAM
OFFICE OF THE DIRECTOR OF MEDICAL EDUCATION, ASSAM
SIXMILE, KHANAPARA, GUWAHATI-22
Tele-Fax No.0361-2366236, E-mail- dmeassam@gmail.com, dme@assam.gov.in,

No. DME/341/2010/ 9851

Dated: 3/09/2014

CORRIGENDUM, ADDENDUM CUM EXTENSION NOTICE

(National Competitive Bidding)

IFB No.DME/341/2010/9452, Dated: 22/08/2014

In modification of tender vide IFB No. referred to above- The Supply & Commissioning of CSSD equipments shall have to be done and quoted on Turnkey basis only. The Para-39 (4) & (5) may be seen.

The technical specification and qualification criteria for CSSD Tender is replaced by the following-

**DETAIL TECHNICAL SPECIFICATIONS OF C.S.S.D. EQUIPMENTS
FOR 500 BEDDED HOSPITAL**

A. Equipments	
1. Horizontal Double Door Steam Sterilizer/Autoclave Pre-vacuum 1000 Liter Capacity Specifications: Fully Automatic CSSD Steam Sterilizer/Autoclave	<ol style="list-style-type: none">1. CSSD Steam Autoclave should be able to be able to sterilize wrapped instruments, unwrapped instruments, linen, glassware and liquids.2. The autoclave should designed to operate on various pre select programs such as Pre- Vacuum Cycle, Gravity Cycle, Porous Load Cycle, Liquid Cycle and two standard Bowie Dick Test and Vacuum Leak Cycle. Autoclave should work up to 136 degree C temp.3. It should be provided with a vertical sliding door, a trolley, a carriage, a steam generator and a dedicated air compressor.4. Sterilizer should meet all international standards such as CE, UL and FDA and both sterilizer as well as the steam generator must have individual ASME Stamping as per pressure vessel code.5. Construction :<ul style="list-style-type: none">• Jacket should be constructed of heavy duty 304L grade stainless steel.• Door should be constructed of heavy duty 304L grade stainless steel• Chamber constructed of heavy duty 316L Ti grade stainless steel.• All the pipes and fittings are made of Stainless Steel and Brass.6. Autoclave should be provided with safety features such as

