

**GOVERNMENT OF ASSAM**  
**OFFICE OF THE DIRECTOR OF MEDICAL EDUCATION, ASSAM**  
**SIXMILE, KHANAPARA, GUWAHATI-22**

No .DME/MRI/241/2012/ 1502

Dated: 28 /01/2013

**INVITATION FOR BID (IFB)**  
**GLOBAL COMPETITIVE BIDDING**  
**IFB NO.DME/MRI/241/2012/15699, DATED. 11-12-2012.**  
**CORRIGENDUM CUM EXTENSION NOTICE**

**SECTION V: TECHNICAL SPECIFICATION (Modified)**


**DETAILED CONFIGURATION FOR A STATE OF THE ART 1.5 TESLA  
 CLINICAL MRI SCANNER FOR F.A.A.M.C&H,BARPETA,ASSAM.**

**Detailed Technical Specification:**

A	B	C
SLNo.	PURCHASER'S REQUIREMENTS	Modifications
<b>5</b>	<b>PATIENT HANDLING SYSTEM</b>	
f	Two way communications should be possible with the patient from the console room.	A second dock able table (MR compatible) to be omitted
<b>6</b>	<b>COMPUTER SYSTEM IMAGE PROCESSOR / OPERATOR CONSOLE</b>	
a	Computer should be latest in the industry, fast and efficient,	
	Largest field of view should be at least <b>45 cm</b> in all three axis instead of <b>48 cm</b> .	The row to be included

The following refinements are incorporated-

1. Turnkey offer should be in compliance with the criteria laid down by H.O.D. Radiology F.A.A.M.C & H, Barpeta.
2. The last date of submission tender/bids extended up **11<sup>th</sup> of February'2013 (Monday)**. The other terms and conditions are remain same.

  
 Director of Medical Education, Assam